

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042819.D
 Acq On : 02 Aug 2024 02:42
 Operator : JC/MD
 Sample : P3410-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-S-864-9-14-073024

Quant Time: Aug 02 05:41:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 08/03/2024
 Supervised By :Semsettin Yesilyurt 08/03/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	57290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	114369	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43486	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51963	56.456	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 112.920%			
35) Dibromofluoromethane	5.385	113	38823	49.138	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.280%			
50) Toluene-d8	8.647	98	134225	46.805	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 93.600%			
62) 4-Bromofluorobenzene	11.079	95	49957	47.668	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.340%			
Target Compounds					Qvalue	
4) Vinyl Chloride	1.374	62	14330	20.166	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	113202	157.362	ug/l	98
12) 1,1-Dichloroethene	2.313	96	14422	20.683	ug/l	88
16) Acetone	2.386	43	4971	14.874	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	31430m	43.685	ug/l	
24) 1,1-Dichloroethane	3.617	63	41435	33.192	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	5489059	6526.556	ug/l	81
30) Chloroform	5.099	83	2490	1.794	ug/l	91
32) 1,1,1-Trichloroethane	5.379	97	13896	11.271	ug/l	97
40) Benzene	6.044	78	6632	2.063	ug/l	99
42) 1,2-Dichloroethane	6.098	62	6529	5.623	ug/l	95
44) Trichloroethene	7.129	130	1162043	1380.648	ug/l	97
52) Toluene	8.726	92	2996	1.436	ug/l	95
55) 1,1,2-Trichloroethane	9.159	97	3572	4.234	ug/l	95
64) Tetrachloroethene	9.275	164	7189	9.837	ug/l	96
65) Chlorobenzene	10.079	112	1836	0.771	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042819.D
Acq On : 02 Aug 2024 02:42
Operator : JC/MD
Sample : P3410-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-S-864-9-14-073024

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/03/2024
Supervised By :Semsettin Yesilyurt 08/03/2024

Quant Time: Aug 02 05:41:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

