

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042826.D
 Acq On : 02 Aug 2024 05:25
 Operator : JC/MD
 Sample : P3430-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 925-K1-WPO-0-0.25-073124

Quant Time: Aug 02 08:29:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	79643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	156317	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	152527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64710	50.573	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.140%
35) Dibromofluoromethane	5.385	113	53055	49.132	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.260%
50) Toluene-d8	8.647	98	189751	48.411	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.820%
62) 4-Bromofluorobenzene	11.079	95	71388	49.838	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.680%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	321351	325.304	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	154218	154.210	ug/l	99
12) 1,1-Dichloroethene	2.319	96	9307	9.601	ug/l #	70
16) Acetone	2.392	43	2509	5.400	ug/l	91
21) trans-1,2-Dichloroethene	3.093	96	21926	21.922	ug/l	84
24) 1,1-Dichloroethane	3.605	63	21191	12.211	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	3067376	2623.520	ug/l	81
32) 1,1,1-Trichloroethane	5.373	97	2990	1.745	ug/l #	93
40) Benzene	6.043	78	5083	1.157	ug/l	93
44) Trichloroethene	7.129	130	54701	47.551	ug/l	93
52) Toluene	8.720	92	7852	2.754	ug/l	98
64) Tetrachloroethene	9.275	164	1901	1.840	ug/l	96
65) Chlorobenzene	10.085	112	1976	0.587	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042826.D
Acq On : 02 Aug 2024 05:25
Operator : JC/MD
Sample : P3430-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
925-K1-WPO-0-0.25-073124

Quant Time: Aug 02 08:29:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

