

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042828.D
 Acq On : 02 Aug 2024 06:12
 Operator : JC/MD
 Sample : P3429-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 931-K1-WS-073124

Quant Time: Aug 02 08:30:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124308	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	53252	52.295	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.580%
35) Dibromofluoromethane	5.391	113	42607	49.616	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.240%
50) Toluene-d8	8.647	98	149519	47.969	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.940%
62) 4-Bromofluorobenzene	11.079	95	53390	46.871	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.740%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2933	7.932	ug/l	# 73
27) cis-1,2-Dichloroethene	4.495	96	2531	2.720	ug/l	76
52) Toluene	8.720	92	6764	2.983	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042828.D
Acq On : 02 Aug 2024 06:12
Operator : JC/MD
Sample : P3429-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
931-K1-WS-073124

Quant Time: Aug 02 08:30:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

