

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042829.D
 Acq On : 02 Aug 2024 06:35
 Operator : JC/MD
 Sample : P3402-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-222

Quant Time: Aug 02 08:31:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55813	51.396	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.800%
35) Dibromofluoromethane	5.379	113	44465	48.354	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.700%
50) Toluene-d8	8.647	98	162229	48.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.079	95	60318	49.449	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.900%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	29613	75.098	ug/l #	89
20) Methylene Chloride	2.788	84	4770	4.521	ug/l	89
25) 2-Butanone	4.587	43	3002	5.182	ug/l	98
43) Isopropyl Acetate	6.342	43	55142	28.960	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042829.D
Acq On : 02 Aug 2024 06:35
Operator : JC/MD
Sample : P3402-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-222

Quant Time: Aug 02 08:31:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

