

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042830.D
 Acq On : 02 Aug 2024 06:58
 Operator : JC/MD
 Sample : P3402-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 355

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/03/2024
 Supervised By :Semsettin Yesilyurt 08/03/2024

Quant Time: Aug 02 08:31:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55774	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60427	52.234	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.460%			
35) Dibromofluoromethane	5.385	113	47874	47.863	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.720%			
50) Toluene-d8	8.647	98	179790	49.521	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.040%			
62) 4-Bromofluorobenzene	11.079	95	63926	48.181	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.360%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	27608	65.722	ug/l	100
18) Methyl Acetate	2.709	43	1429	1.355	ug/l	94
20) Methylene Chloride	2.788	84	5228	4.675	ug/l	95
25) 2-Butanone	4.599	43	3314m	5.369	ug/l	
43) Isopropyl Acetate	6.348	43	55313	26.707	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042830.D
Acq On : 02 Aug 2024 06:58
Operator : JC/MD
Sample : P3402-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
355

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/03/2024
Supervised By :Semsettin Yesilyurt 08/03/2024

Quant Time: Aug 02 08:31:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

