

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042832.D
 Acq On : 02 Aug 2024 07:45
 Operator : JC/MD
 Sample : PB162393ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB162393ZHE#01

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda

Quant Time: Aug 02 09:20:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	65945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	136300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132996	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60180	56.802	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.600%
35) Dibromofluoromethane	5.385	113	47108	50.031	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	8.647	98	161795	47.341	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.680%
62) 4-Bromofluorobenzene	11.079	95	61428	49.183	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.360%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	25341	65.871	ug/l #	81
18) Methyl Acetate	2.715	43	2333	2.416	ug/l #	86
20) Methylene Chloride	2.794	84	5649	5.664	ug/l #	75
25) 2-Butanone	4.587	43	3280m	5.803	ug/l	
43) Isopropyl Acetate	6.342	43	62452	32.033	ug/l #	95

08/03/2024
 Supervised By :Semsettin
 Kesilyurt

08/03/2024

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042832.D
Acq On : 02 Aug 2024 07:45
Operator : JC/MD
Sample : PB162393ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB162393ZHE#01

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda

Quant Time: Aug 02 09:20:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

