

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080125\
 Data File : VX047207.D
 Acq On : 01 Aug 2025 15:41
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 02 00:26:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	223902	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	403808	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	182165	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	160917	57.695	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.380%
35) Dibromofluoromethane	5.397	113	134080	53.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.560%
50) Toluene-d8	8.653	98	379048	46.945	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.880%
62) 4-Bromofluorobenzene	11.079	95	179615	51.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080125\
Data File : VX047207.D
Acq On : 01 Aug 2025 15:41
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 02 00:26:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

