

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080125\
 Data File : VX047202.D
 Acq On : 01 Aug 2025 10:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 00:22:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.664	0.590	11.1	62	0.00
3 P	Chloromethane	0.657	0.535	18.6	62	0.00
4 C	Vinyl Chloride	0.706	0.648	8.2#	67	0.00
5 T	Bromomethane	0.384	0.348	9.4	63	0.01
6 T	Chloroethane	0.464	0.382	17.7	65	0.00
7 T	Trichlorofluoromethane	1.053	0.974	7.5	68	0.00
8 T	Diethyl Ether	0.371	0.389	-4.9	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.634	0.578	8.8	69	0.00
10 T	Methyl Iodide	0.647	0.918	-41.9#	101	0.00
11 T	Tert butyl alcohol	0.080	0.069	13.7	75	0.00
12 CM	1,1-Dichloroethene	0.628	0.571	9.1#	68	0.00
13 T	Acrolein	0.134	0.132	1.5	77	0.00
14 T	Allyl chloride	1.136	1.020	10.2	66	0.00
15 T	Acrylonitrile	0.297	0.308	-3.7	75	0.00
16 T	Acetone	0.250	0.273	-9.2	87	0.00
17 T	Carbon Disulfide	1.818	1.444	20.6	60	0.00
18 T	Methyl Acetate	0.663	0.824	-24.3	96	0.00
19 T	Methyl tert-butyl Ether	1.912	2.076	-8.6	81	0.00
20 T	Methylene Chloride	0.686	0.683	0.4	76	0.00
21 T	trans-1,2-Dichloroethene	0.643	0.601	6.5	70	0.00
22 T	Diisopropyl ether	2.184	2.335	-6.9	80	0.00
23 T	Vinyl Acetate	1.752	1.882	-7.4	78	0.00
24 P	1,1-Dichloroethane	1.214	1.215	-0.1	75	0.00
25 T	2-Butanone	0.363	0.371	-2.2	76	0.00
26 T	2,2-Dichloropropane	0.940	0.915	2.7	75	0.00
27 T	cis-1,2-Dichloroethene	0.761	0.764	-0.4	76	0.00
28 T	Bromochloromethane	0.559	0.657	-17.5	87	0.00
29 T	Tetrahydrofuran	0.234	0.221	5.6	69	0.00
30 C	Chloroform	1.236	1.249	-1.1#	78	0.00
31 T	Cyclohexane	1.164	0.965	17.1	64	0.00
32 T	1,1,1-Trichloroethane	1.074	1.013	5.7	72	-0.01
33 S	1,2-Dichloroethane-d4	0.623	0.677	-8.7	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.309	0.325	-5.2	94	0.00
36 T	1,1-Dichloropropene	0.526	0.437	16.9	68	0.00
37 T	Ethyl Acetate	0.485	0.477	1.6	79	0.00
38 T	Carbon Tetrachloride	0.562	0.496	11.7	71	-0.01
39 T	Methylcyclohexane	0.657	0.545	17.0	66	0.00
40 TM	Benzene	1.523	1.415	7.1	73	0.00
41 T	Methacrylonitrile	0.254	0.247	2.8	72	0.00
42 TM	1,2-Dichloroethane	0.536	0.541	-0.9	80	-0.01
43 T	Isopropyl Acetate	0.756	0.773	-2.2	77	-0.01
44 TM	Trichloroethene	0.391	0.346	11.5	70	0.00
45 C	1,2-Dichloropropane	0.381	0.375	1.6#	78	0.00
46 T	Dibromomethane	0.267	0.271	-1.5	78	0.00
47 T	Bromodichloromethane	0.569	0.581	-2.1	80	0.00
48 T	Methyl methacrylate	0.412	0.402	2.4	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080125\
 Data File : VX047202.D
 Acq On : 01 Aug 2025 10:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 00:22:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.004	20.0	68	0.06
50 S	Toluene-d8	1.000	0.928	7.2	85	0.00
51 T	4-Methyl-2-Pentanone	0.449	0.446	0.7	76	0.00
52 CM	Toluene	0.936	0.871	6.9#	73	0.00
53 T	t-1,3-Dichloropropene	0.538	0.551	-2.4	79	0.00
54 T	cis-1,3-Dichloropropene	0.588	0.591	-0.5	77	0.00
55 T	1,1,2-Trichloroethane	0.351	0.359	-2.3	81	0.00
56 T	Ethyl methacrylate	0.520	0.545	-4.8	78	0.00
57 T	1,3-Dichloropropane	0.610	0.610	0.0	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.289	0.295	-2.1	78	0.00
59 T	2-Hexanone	0.297	0.300	-1.0	74	0.00
60 T	Dibromochloromethane	0.412	0.428	-3.9	82	0.00
61 T	1,2-Dibromoethane	0.367	0.366	0.3	79	0.00
62 S	4-Bromofluorobenzene	0.431	0.443	-2.8	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.360	0.305	15.3	69	0.00
65 PM	Chlorobenzene	1.171	1.093	6.7	76	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.390	1.0	79	0.00
67 C	Ethyl Benzene	2.048	1.881	8.2#	73	0.00
68 T	m/p-Xylenes	0.766	0.696	9.1	73	0.00
69 T	o-Xylene	0.729	0.691	5.2	75	0.00
70 T	Styrene	1.250	1.216	2.7	76	0.00
71 P	Bromoform	0.300	0.300	0.0	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.945	3.600	8.7	73	0.00
74 T	N-amyl acetate	1.592	1.540	3.3	75	0.00
75 P	1,1,2,2-Tetrachloroethane	1.127	1.112	1.3	80	0.00
76 T	1,2,3-Trichloropropane	0.925	0.914	1.2	80	0.00
77 T	Bromobenzene	0.946	0.891	5.8	76	0.00
78 T	n-propylbenzene	4.714	4.274	9.3	73	0.00
79 T	2-Chlorotoluene	2.818	2.596	7.9	75	0.00
80 T	1,3,5-Trimethylbenzene	3.259	3.005	7.8	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.241	32.9#	54	0.00
82 T	4-Chlorotoluene	3.289	3.085	6.2	75	0.00
83 T	tert-Butylbenzene	3.271	3.067	6.2	75	0.00
84 T	1,2,4-Trimethylbenzene	3.251	3.032	6.7	74	0.00
85 T	sec-Butylbenzene	4.087	3.710	9.2	74	0.00
86 T	p-Isopropyltoluene	3.389	3.126	7.8	74	0.00
87 T	1,3-Dichlorobenzene	1.798	1.665	7.4	76	0.00
88 T	1,4-Dichlorobenzene	1.821	1.678	7.9	77	0.00
89 T	n-Butylbenzene	3.178	2.897	8.8	73	0.00
90 T	Hexachloroethane	0.606	0.567	6.4	75	0.00
91 T	1,2-Dichlorobenzene	1.718	1.608	6.4	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.216	0.218	-0.9	79	0.00
93 T	1,2,4-Trichlorobenzene	1.135	1.108	2.4	79	0.00
94 T	Hexachlorobutadiene	0.386	0.396	-2.6	85	0.00
95 T	Naphthalene	3.328	3.344	-0.5	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080125\
Data File : VX047202.D
Acq On : 01 Aug 2025 10:22
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 02 00:22:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.094	1.072	2.0	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6