

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080218\
 Data File : VX003819.D
 Acq On : 02 Aug 2018 12:03
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Aug 03 03:25:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X072718W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jul 27 16:17:43 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	99	0.00
2 M	Dichlorodifluoromethane	1.585	2.085	-31.5#	144	0.00
3 M	Chloromethane	2.133	2.572	-20.6	132	0.00
4 M	Vinyl Chloride	2.243	1.964	12.4	96	0.00
5 M	Bromomethane	1.341	1.045	22.1	91	0.00
6 M	Chloroethane	1.468	1.240	15.5	92	0.00
7 M	Trichlorofluoromethane	3.017	2.633	12.7	97	0.00
8 T	Diethyl Ether	1.293	1.083	16.2	93	0.00
9	1,1,2-Trichlorotrifluoroeth	1.829	1.612	11.9	96	0.00
10 M	1,1-Dichloroethene	1.720	1.570	8.7	101	0.00
11	Methyl Iodide	2.384	1.654	30.6#	77	0.00
12	Methyl Acetate	3.182	0.958	69.9#	32	0.43
13 M	Acrolein	0.248	0.293	-18.1	114	0.00
14 M	Acrylonitrile	1.254	0.991	21.0	85	0.00
15 M	Acetone	0.340	0.293	13.8	93	0.00
16 M	Carbon Disulfide	4.776	4.169	12.7	98	0.00
17	Allyl chloride	3.560	2.861	19.6	88	0.00
18 M	Methylene Chloride	2.037	1.854	9.0	98	0.00
19 M	trans-1,2-Dichloroethene	1.859	1.565	15.8	91	0.00
20 T	Diisopropyl ether	6.537	5.420	17.1	89	0.00
21 M	1,1-Dichloroethane	3.701	3.110	16.0	91	0.00
22 M	cis-1,2-Dichloroethene	1.877	1.926	-2.6	105	0.00
23 M	tert-Butyl Alcohol	0.527	0.416	21.1	84	-0.01
24 M	Methyl tert-Butyl Ether	6.169	5.218	15.4	93	-0.01
25 M	Chloroform	3.205	3.103	3.2	98	0.00
26	Cyclohexane	2.780	2.673	3.8	100	-0.01
27 s	1,2-Dichloroethane-d4	2.202	2.001	9.1	95	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
29	1,1-Dichloropropene	0.443	0.460	-3.8	97	0.00
30 M	2-Butanone	0.283	0.294	-3.9	93	-0.01
31	2,2-Dichloropropane	0.429	0.489	-14.0	104	-0.01
32 M	1,1,1-Trichloroethane	0.507	0.522	-3.0	96	0.00
33 M	Carbon Tetrachloride	0.440	0.464	-5.5	97	-0.01
34 M	Benzene	1.361	1.433	-5.3	96	-0.01
35	Methacrylonitrile	0.294	0.312	-6.1	98	0.00
36 M	1,2-Dichloroethane	0.489	0.478	2.2	91	0.00
37 M	Trichloroethene	0.355	0.357	-0.6	93	0.00
38	Methylcyclohexane	0.503	0.494	1.8	93	0.00
39 M	1,2-Dichloropropane	0.360	0.357	0.8	92	0.00
40	Dibromomethane	0.233	0.233	0.0	92	0.00
41 M	Bromodichloromethane	0.445	0.429	3.6	92	0.00
42 M	Vinyl Acetate	1.087	0.913	16.0	84	-0.01
43	Ethyl Acetate	0.528	0.566	-7.2	97	0.00
44	Isopropyl Acetate	0.811	0.760	6.3	88	0.00
45 T	1,4-Dioxane	0.009	0.009#	0.0	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080218\
 Data File : VX003819.D
 Acq On : 02 Aug 2018 12:03
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Aug 03 03:25:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X072718W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jul 27 16:17:43 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.418	0.398	4.8	89	0.00
47	n-amyl Acetate	0.765	0.672	12.2	89	0.00
48 M	t-1,3-Dichloropropene	0.540	0.537	0.6	95	0.00
49 T	cis-1,3-Dichloropropene	0.506	0.488	3.6	98	0.00
50 M	1,1,2-Trichloroethane	0.370	0.357	3.5	94	0.00
51	Ethyl methacrylate	0.556	0.501	9.9	92	0.00
52	1,3-Dichloropropane	0.615	0.590	4.1	94	0.00
53 M	Dibromochloromethane	0.363	0.353	2.8	98	0.00
54 M	1,2-Dibromoethane	0.386	0.372	3.6	95	0.00
55 M	2-Chloroethyl vinyl ether	0.287	0.264	8.0	88	0.00
56 M	Bromoform	0.282	0.269	4.6	100	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
58 M	4-Methyl-2-Pentanone	0.566	0.549	3.0	86	0.00
59 M	2-Hexanone	0.451	0.417	7.5	86	0.00
60 S	4-Bromofluorobenzene	0.483	0.485	-0.4	96	0.00
61 M	Tetrachloroethene	0.349	0.362	-3.7	100	0.00
62 M	Toluene	1.557	1.559	-0.1	92	0.00
63 S	Toluene-d8	1.351	1.339	0.9	90	0.00
64 M	Chlorobenzene	1.006	1.017	-1.1	94	0.00
65	1,1,1,2-Tetrachloroethane	0.356	0.364	-2.2	96	0.00
66 M	Ethyl Benzene	1.713	1.680	1.9	93	0.00
67 M	m/p-Xylenes	0.660	0.666	-0.9	95	0.00
68 M	o-Xylene	0.631	0.635	-0.6	95	0.00
69 M	Styrene	1.069	1.061	0.7	94	0.00
70	Isopropylbenzene	1.738	1.701	2.1	94	0.00
71 M	1,1,2,2-Tetrachloroethane	0.621	0.595	4.2	91	0.00
72	1,2,3-Trichloropropane	0.516	0.538	-4.3	99	0.00
73	Bromobenzene	0.468	0.477	-1.9	100	0.00
74	n-propylbenzene	2.048	2.002	2.2	94	0.00
75	2-Chlorotoluene	1.207	1.197	0.8	94	0.00
76	1,3,5-Trimethylbenzene	1.475	1.454	1.4	95	0.00
77	t-1,4-Dichloro-2-butene	0.176	0.171	2.8	98	0.00
78	4-Chlorotoluene	1.412	1.395	1.2	95	0.00
79	tert-butylbenzene	1.518	1.517	0.1	96	0.00
80	1,2,4-Trimethylbenzene	1.510	1.493	1.1	94	0.00
81	sec-Butylbenzene	1.718	1.699	1.1	96	0.00
82	p-Isopropyltoluene	1.518	1.517	0.1	96	0.00
83 M	1,3-Dichlorobenzene	0.856	0.886	-3.5	99	0.00
84 M	1,4-Dichlorobenzene	0.861	0.903	-4.9	100	0.00
85	n-Butylbenzene	1.347	1.320	2.0	98	0.00
86 T	Hexachloroethane	0.245	0.224	8.6	90	0.00
87 M	1,2-Dichlorobenzene	0.859	0.889	-3.5	99	0.00
88	1,2-Dibromo-3-Chloropropane	0.131	0.124	5.3	95	0.00
89	1,2,4-Trichlorobenzene	0.603	0.661	-9.6	108	0.00
90	Hexachlorobutadiene	0.273	0.304	-11.4	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080218\
Data File : VX003819.D
Acq On : 02 Aug 2018 12:03
Operator : JC/MD
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Aug 03 03:25:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X072718W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jul 27 16:17:43 2018
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.843	1.921	-4.2	100	0.00
92	1,2,3-Trichlorobenzene	0.614	0.675	-9.9	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0