

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011299.D
 Acq On : 02 Aug 2019 13:00
 Operator : JC/SP
 Sample : K4119-01 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0683

Quant Time: Aug 03 06:35:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	136609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	217115	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	205158	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	96522	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	71136	51.71	ug/l	0.00
Spiked Amount						
			Recovery	=	103.42%	
35) Dibromofluoromethane	5.49	113	64721	45.66	ug/l	0.00
Spiked Amount						
			Recovery	=	91.32%	
50) Toluene-d8	8.71	98	262137	48.99	ug/l	0.00
Spiked Amount						
			Recovery	=	97.98%	
62) 4-Bromofluorobenzene	11.13	95	88313	46.23	ug/l	0.00
Spiked Amount						
			Recovery	=	92.46%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011299.D
Acq On : 02 Aug 2019 13:00
Operator : JC/SP
Sample : K4119-01 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0683

Quant Time: Aug 03 06:35:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

