

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011300.D
 Acq On : 02 Aug 2019 13:23
 Operator : JC/SP
 Sample : K4119-02 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0682

Quant Time: Aug 03 06:38:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	149766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	238491	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	226105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	103822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	77640	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
35) Dibromofluoromethane	5.49	113	73187	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
50) Toluene-d8	8.71	98	288641	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
62) 4-Bromofluorobenzene	11.13	95	95570	45.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.08%	
Target Compounds						
6) Chloroethane	1.71	64	107	Below Cal	#	10
27) cis-1,2-Dichloroethene	4.59	96	1760	1.060	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011300.D
Acq On : 02 Aug 2019 13:23
Operator : JC/SP
Sample : K4119-02 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0682

Quant Time: Aug 03 06:38:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

