

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011303.D
 Acq On : 02 Aug 2019 14:33
 Operator : JC/SP
 Sample : K4119-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-TB-2019-4

Quant Time: Aug 03 06:44:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	143733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	234235	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	218858	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	100030	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	77593	53.61	ug/l	0.00
Spiked Amount						
			Recovery	=	107.22%	
35) Dibromofluoromethane	5.49	113	72546	47.44	ug/l	0.00
Spiked Amount						
			Recovery	=	94.88%	
50) Toluene-d8	8.71	98	277765	48.12	ug/l	0.00
Spiked Amount						
			Recovery	=	96.24%	
62) 4-Bromofluorobenzene	11.13	95	92798	45.03	ug/l	0.00
Spiked Amount						
			Recovery	=	90.06%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011303.D
Acq On : 02 Aug 2019 14:33
Operator : JC/SP
Sample : K4119-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-TB-2019-4

Quant Time: Aug 03 06:44:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

