

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011306.D
 Acq On : 02 Aug 2019 15:43
 Operator : JC/SP
 Sample : K4119-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0682

Quant Time: Aug 03 06:51:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	142739	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	227165	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	217680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	98961	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	75404	52.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.92%	
35) Dibromofluoromethane	5.49	113	72526	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
50) Toluene-d8	8.71	98	277701	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
62) 4-Bromofluorobenzene	11.13	95	93716	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	6910	5.074	ug/l #	86
16) Acetone	2.44	43	1578	2.306	ug/l	99
21) trans-1,2-Dichloroethene	3.16	96	1391	1.013	ug/l #	83
27) cis-1,2-Dichloroethene	4.59	96	30275	19.133	ug/l	90
40) Benzene	6.13	78	2473	0.418	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011306.D
Acq On : 02 Aug 2019 15:43
Operator : JC/SP
Sample : K4119-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0682

Quant Time: Aug 03 06:51:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

