

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011311.D
 Acq On : 02 Aug 2019 17:40
 Operator : JC/SP
 Sample : K4119-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0678

Quant Time: Aug 03 06:59:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	142789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	228826	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	217874	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	99686	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	76002	52.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.70%	
35) Dibromofluoromethane	5.49	113	73609	49.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.54%	
50) Toluene-d8	8.71	98	281069	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
62) 4-Bromofluorobenzene	11.13	95	94046	46.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.42%	
Target Compounds						
16) Acetone	2.44	43	1790	2.615	ug/l #	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011311.D
Acq On : 02 Aug 2019 17:40
Operator : JC/SP
Sample : K4119-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0678

Quant Time: Aug 03 06:59:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

