

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011316.D
 Acq On : 02 Aug 2019 19:37
 Operator : JC/SP
 Sample : K4119-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0689

Quant Time: Aug 03 07:14:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	137458	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	224213	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	209760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	94406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	76196	55.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.08%	
35) Dibromofluoromethane	5.49	113	72557	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
50) Toluene-d8	8.71	98	271455	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
62) 4-Bromofluorobenzene	11.13	95	91270	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3991	3.043	ug/l	91
21) trans-1,2-Dichloroethene	3.17	96	2811	2.125	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	12062	7.916	ug/l	94
65) Chlorobenzene	10.13	112	13865	3.083	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011316.D
Acq On : 02 Aug 2019 19:37
Operator : JC/SP
Sample : K4119-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0689

Quant Time: Aug 03 07:14:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

