

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011317.D
 Acq On : 02 Aug 2019 20:01
 Operator : JC/SP
 Sample : K4119-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0690

Quant Time: Aug 03 07:18:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	138133	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	223420	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	212329	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95855	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	74875	53.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.66%	
35) Dibromofluoromethane	5.49	113	72231	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
50) Toluene-d8	8.71	98	270018	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
62) 4-Bromofluorobenzene	11.13	95	91611	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
Target Compounds						
4) Vinyl Chloride	1.40	62	11316	8.587	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	1944	1.269	ug/l	97
40) Benzene	6.13	78	3765	0.647	ug/l	98
65) Chlorobenzene	10.13	112	32357	7.107	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011317.D
Acq On : 02 Aug 2019 20:01
Operator : JC/SP
Sample : K4119-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0690

Quant Time: Aug 03 07:18:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

