

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011318.D
 Acq On : 02 Aug 2019 20:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 07:19:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.386	0.358	7.3	96	0.00
3 P	Chloromethane	0.434	0.394	9.2	95	0.00
4 C	Vinyl Chloride	0.477	0.453	5.0#	97	0.00
5 T	Bromomethane	0.340	0.253	25.6#	77	0.00
6 T	Chloroethane	0.335	0.313	6.6	101	0.00
7 T	Trichlorofluoromethane	0.848	0.825	2.7	97	0.00
8 T	Diethyl Ether	0.303	0.292	3.6	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.459	0.431	6.1	96	0.00
10 T	Methyl Iodide	0.579	0.385	33.5#	62	0.00
11 T	Tert butyl alcohol	0.106	0.105	0.9	103	0.00
12 CM	1,1-Dichloroethene	0.455	0.436	4.2#	99	0.00
13 T	Acrolein	0.066	0.052	21.2#	86	0.00
14 T	Allyl chloride	0.606	0.619	-2.1	99	0.00
15 T	Acrylonitrile	0.240	0.249	-3.8	102	0.00
16 T	Acetone	0.240	0.203	15.4	88	0.00
17 T	Carbon Disulfide	1.068	1.038	2.8	96	0.00
18 T	Methyl Acetate	0.548	0.561	-2.4	106	0.00
19 T	Methyl tert-butyl Ether	1.341	1.372	-2.3	102	0.00
20 T	Methylene Chloride	0.525	0.503	4.2	102	0.00
21 T	trans-1,2-Dichloroethene	0.481	0.474	1.5	99	0.00
22 T	Diisopropyl ether	1.333	1.356	-1.7	104	0.00
23 T	Vinyl Acetate	1.106	1.193	-7.9	102	0.00
24 P	1,1-Dichloroethane	0.786	0.781	0.6	101	0.00
25 T	2-Butanone	0.341	0.341	0.0	99	0.00
26 T	2,2-Dichloropropane	0.611	0.539	11.8	87	0.00
27 T	cis-1,2-Dichloroethene	0.554	0.540	2.5	101	0.00
28 T	Bromochloromethane	0.358	0.352	1.7	105	0.00
29 T	Tetrahydrofuran	0.209	0.220	-5.3	103	0.00
30 C	Chloroform	0.854	0.853	0.1#	103	0.00
31 T	Cyclohexane	0.648	0.629	2.9	97	0.00
32 T	1,1,1-Trichloroethane	0.732	0.732	0.0	100	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.507	-0.6	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.326	0.313	4.0	105	0.00
36 T	1,1-Dichloropropene	0.437	0.398	8.9	97	0.00
37 T	Ethyl Acetate	0.438	0.426	2.7	103	0.00
38 T	Carbon Tetrachloride	0.439	0.428	2.5	100	0.00
39 T	Methylcyclohexane	0.517	0.457	11.6	94	0.00
40 TM	Benzene	1.303	1.253	3.8	100	0.00
41 T	Methacrylonitrile	0.239	0.233	2.5	101	0.00
42 TM	1,2-Dichloroethane	0.441	0.426	3.4	104	0.00
43 T	Isopropyl Acetate	0.700	0.687	1.9	103	0.00
44 TM	Trichloroethene	0.394	0.362	8.1	98	0.00
45 C	1,2-Dichloropropane	0.337	0.318	5.6#	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
 Data File : VX011318.D
 Acq On : 02 Aug 2019 20:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 07:19:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.246	0.233	5.3	102	0.00
47 T	Bromodichloromethane	0.417	0.412	1.2	101	0.00
48 T	Methyl methacrylate	0.329	0.333	-1.2	105	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	98	0.00
50 S	Toluene-d8	1.232	1.178	4.4	105	0.00
51 T	4-Methyl-2-Pentanone	0.452	0.472	-4.4	105	0.00
52 CM	Toluene	0.861	0.829	3.7#	100	0.00
53 T	t-1,3-Dichloropropene	0.423	0.437	-3.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.487	0.484	0.6	97	0.00
55 T	1,1,2-Trichloroethane	0.354	0.350	1.1	103	0.00
56 T	Ethyl methacrylate	0.505	0.514	-1.8	103	0.00
57 T	1,3-Dichloropropane	0.560	0.549	2.0	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.248	0.235	5.2	97	0.00
59 T	2-Hexanone	0.350	0.371	-6.0	104	0.00
60 T	Dibromochloromethane	0.363	0.376	-3.6	104	0.00
61 T	1,2-Dibromoethane	0.383	0.375	2.1	101	0.00
62 S	4-Bromofluorobenzene	0.440	0.431	2.0	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.425	0.386	9.2	100	0.00
65 PM	Chlorobenzene	1.072	1.021	4.8	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.371	-1.1	102	0.00
67 C	Ethyl Benzene	1.795	1.741	3.0#	102	0.00
68 T	m/p-Xylenes	0.694	0.676	2.6	100	0.00
69 T	o-Xylene	0.682	0.658	3.5	101	0.00
70 T	Styrene	1.129	1.141	-1.1	101	0.00
71 P	Bromoform	0.285	0.294	-3.2	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	3.420	3.279	4.1	101	0.00
74 T	N-amyl acetate	1.236	1.287	-4.1	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.140	1.117	2.0	103	0.00
76 T	1,2,3-Trichloropropane	0.911	0.923	-1.3	104	0.00
77 T	Bromobenzene	0.948	0.897	5.4	100	0.00
78 T	n-propylbenzene	3.803	3.653	3.9	100	0.00
79 T	2-Chlorotoluene	2.267	2.151	5.1	101	0.00
80 T	1,3,5-Trimethylbenzene	2.863	2.794	2.4	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.312	0.295	5.4	94	0.00
82 T	4-Chlorotoluene	2.642	2.541	3.8	102	0.00
83 T	tert-Butylbenzene	2.808	2.714	3.3	100	0.00
84 T	1,2,4-Trimethylbenzene	2.850	2.792	2.0	100	0.00
85 T	sec-Butylbenzene	3.373	3.250	3.6	101	0.00
86 T	p-Isopropyltoluene	3.103	2.991	3.6	99	0.00
87 T	1,3-Dichlorobenzene	1.675	1.606	4.1	100	0.00
88 T	1,4-Dichlorobenzene	1.728	1.590	8.0	98	0.00
89 T	n-Butylbenzene	2.674	2.508	6.2	96	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080219\
Data File : VX011318.D
Acq On : 02 Aug 2019 20:24
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 03 07:19:32 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.459	0.450	2.0	96	0.00
91 T	1,2-Dichlorobenzene	1.683	1.605	4.6	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.221	0.223	-0.9	101	0.00
93 T	1,2,4-Trichlorobenzene	1.116	1.045	6.4	96	0.00
94 T	Hexachlorobutadiene	0.566	0.520	8.1	96	0.00
95 T	Naphthalene	3.265	3.366	-3.1	102	0.00
96 T	1,2,3-Trichlorobenzene	1.135	1.077	5.1	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6