

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080222\
 Data File : VX030342.D
 Acq On : 02 Aug 2022 10:53
 Operator : JC/MD
 Sample : VX0802WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0802WBL01

Quant Time: Aug 02 15:48:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	251360	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139998	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163351	55.940	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.880%
35) Dibromofluoromethane	5.391	113	129483	53.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.780%
50) Toluene-d8	8.653	98	480638	51.651	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.300%
62) 4-Bromofluorobenzene	11.079	95	164818	47.725	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080222\
Data File : VX030342.D
Acq On : 02 Aug 2022 10:53
Operator : JC/MD
Sample : VX0802WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0802WBL01

Quant Time: Aug 02 15:48:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

