

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080222\
 Data File : VX030351.D
 Acq On : 02 Aug 2022 14:44
 Operator : JC/MD
 Sample : N3861-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-072222

Quant Time: Aug 03 01:07:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	152764	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	262787	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	148509	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	149509	48.693	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.380%
35) Dibromofluoromethane	5.391	113	120714	48.053	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.653	98	461362	47.424	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.840%
62) 4-Bromofluorobenzene	11.085	95	161466	44.721	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.440%
Target Compounds						
95) Naphthalene	13.780	128	5933	0.504	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080222\
Data File : VX030351.D
Acq On : 02 Aug 2022 14:44
Operator : JC/MD
Sample : N3861-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-072222

Quant Time: Aug 03 01:07:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

