

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080223\
 Data File : VX036763.D
 Acq On : 02 Aug 2023 12:16
 Operator : JC/MD
 Sample : VX0802WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0802WBL01

Quant Time: Aug 03 07:27:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	265941	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105965	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100504	46.085	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.160%
35) Dibromofluoromethane	5.385	113	84612	46.288	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.580%
50) Toluene-d8	8.647	98	316897	48.887	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.780%
62) 4-Bromofluorobenzene	11.079	95	115512	47.232	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080223\
Data File : VX036763.D
Acq On : 02 Aug 2023 12:16
Operator : JC/MD
Sample : VX0802WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0802WBL01

Quant Time: Aug 03 07:27:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

