

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080223\
 Data File : VX036768.D
 Acq On : 02 Aug 2023 14:17
 Operator : JC/MD
 Sample : 03862-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK-1

Quant Time: Aug 03 07:28:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	259588	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	238962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98654	46.430	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.860%		
35) Dibromofluoromethane	5.391	113	83308	46.690	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.380%		
50) Toluene-d8	8.647	98	312736	49.426	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.860%		
62) 4-Bromofluorobenzene	11.079	95	115207	48.260	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.520%		
Target Compounds						
16) Acetone	2.386	43	2418	2.742	ug/l	99
20) Methylene Chloride	2.782	84	770	0.393	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080223\
Data File : VX036768.D
Acq On : 02 Aug 2023 14:17
Operator : JC/MD
Sample : 03862-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK-1

Quant Time: Aug 03 07:28:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

