

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080223\
 Data File : VX036769.D
 Acq On : 02 Aug 2023 14:41
 Operator : JC/MD
 Sample : 03862-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-4

Quant Time: Aug 03 07:28:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240294	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90232	45.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.820%
35) Dibromofluoromethane	5.391	113	75427	45.668	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.647	98	283412	48.388	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.780%
62) 4-Bromofluorobenzene	11.079	95	103039	46.628	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.260%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1794	2.199	ug/l	100
20) Methylene Chloride	2.788	84	1076	0.593	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080223\
Data File : VX036769.D
Acq On : 02 Aug 2023 14:41
Operator : JC/MD
Sample : 03862-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-4

Quant Time: Aug 03 07:28:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

