

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080224\
 Data File : VX042847.D
 Acq On : 02 Aug 2024 17:20
 Operator : JC/MD
 Sample : VX0802WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0802WBL01

Quant Time: Aug 02 21:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080224W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 02 21:32:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	14876	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	86785	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	81657	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	40806	27.956	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	93.200%
60) 4-Bromofluorobenzene	11.079	95	35902	26.359	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.867%
63) Toluene-d8	8.653	98	104132	28.912	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.367%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080224\
Data File : VX042847.D
Acq On : 02 Aug 2024 17:20
Operator : JC/MD
Sample : VX0802WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0802WBL01

Quant Time: Aug 02 21:35:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080224W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 02 21:32:14 2024
Response via : Initial Calibration

