

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080224\
 Data File : VX042852.D
 Acq On : 02 Aug 2024 19:18
 Operator : JC/MD
 Sample : P3386-01 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EFF-WW

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/05/2024
 Supervised By :Semsettin Yesilyurt 08/05/2024

Quant Time: Aug 02 21:36:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080224W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 02 21:32:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	13784m	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	78976	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	75534	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	36785	27.198	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	90.667%#		
60) 4-Bromofluorobenzene	11.079	95	35925	28.514	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	95.033%		
63) Toluene-d8	8.647	98	96608	28.997	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	96.667%		
Target Compounds						
15) Acetone	2.386	58	18482	97.864	ug/l	98
25) Chloroform	5.093	83	2386m	1.251	ug/l	
62) Toluene	8.726	91	7338	1.922	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080224\
Data File : VX042852.D
Acq On : 02 Aug 2024 19:18
Operator : JC/MD
Sample : P3386-01 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EFF-WW

Quant Time: Aug 02 21:36:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080224W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 02 21:32:14 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/05/2024
Supervised By :Semsettin Yesilyurt 08/05/2024

