

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080320\
 Data File : VX017755.D
 Acq On : 03 Aug 2020 19:11
 Operator : JC/SP
 Sample : L3512-05 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 USED-ANTIFREEZE-1456

Manual Integrations
 APPROVED

MMDadoda
 8/4/2020 9:51:45 AM

Quant Time: Aug 04 07:54:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	623360	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	930545	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	895502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	466973	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	409937	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
35) Dibromofluoromethane	5.47	113	316483	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	8.70	98	1120701	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
62) 4-Bromofluorobenzene	11.12	95	439772	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.01	59	10903	6.461	ug/l	96
16) Acetone	2.42	43	259015	73.494	ug/l	96
40) Benzene	6.11	78	6216	0.243	ug/l	97
52) Toluene	8.77	92	20557	1.253	ug/l	99
67) Ethyl Benzene	10.23	91	34460	1.065	ug/l	99
68) m/p-Xylenes	10.34	106	48444	3.865	ug/l	95
69) o-Xylene	10.68	106	32178	2.669	ug/l	99
73) Isopropylbenzene	11.00	105	18425	0.575	ug/l	96
78) n-propylbenzene	11.34	91	73589	2.037	ug/l	97
80) 1,3,5-Trimethylbenzene	11.49	105	78361	2.865	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	324895	11.943	ug/l	100
85) sec-Butylbenzene	11.93	105	32307	1.060	ug/l	80
86) p-Isopropyltoluene	12.01	119	39921m	1.375	ug/l	
89) n-Butylbenzene	12.37	91	71077	2.747	ug/l #	47
95) Naphthalene	13.82	128	127269	4.127	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080320\
Data File : VX017755.D
Acq On : 03 Aug 2020 19:11
Operator : JC/SP
Sample : L3512-05 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
USED-ANTIFREEZE-1456

Manual Integrations
APPROVED

MMDadoda
8/4/2020 9:51:45 AM

Quant Time: Aug 04 07:54:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

