

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080321\
 Data File : VX023525.D
 Acq On : 03 Aug 2021 14:03
 Operator : JC/MD
 Sample : VX0803WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0803WBL01

Quant Time: Aug 04 04:54:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Aug 03 12:05:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.909	128	37356	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	208488	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	174580	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	87309	30.222	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.733%
60) 4-Bromofluorobenzene	11.085	95	75538	26.417	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.067%
63) Toluene-d8	8.653	98	244723	31.147	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.833%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080321\
Data File : VX023525.D
Acq On : 03 Aug 2021 14:03
Operator : JC/MD
Sample : VX0803WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0803WBL01

Quant Time: Aug 04 04:54:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Aug 03 12:05:42 2021
Response via : Initial Calibration

