



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080321\
 Data File : VX023517.D
 Acq On : 03 Aug 2021 09:35
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Tue Aug 03 12:05:42 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	11855	PASS
75	95	30	60	52.3	32293	PASS
95	95	100	100	100.0	61755	PASS
96	95	5	9	6.7	4165	PASS
173	174	0.00	2	0.9	477	PASS
174	95	50	100	81.4	50261	PASS
175	174	5	9	7.1	3592	PASS
176	174	95	101	99.3	49909	PASS
177	176	5	9	6.1	3066	PASS