

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
 Data File : VX030368.D
 Acq On : 03 Aug 2022 14:13
 Operator : JC/MD
 Sample : N3976-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP3-0729

Quant Time: Aug 03 15:55:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180330	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	121357	63.328	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 126.660%#	
35) Dibromofluoromethane	5.397	113	91099	52.846	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.700%	
50) Toluene-d8	8.653	98	321120	48.101	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.200%	
62) 4-Bromofluorobenzene	11.085	95	111487	44.998	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.000%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	72401	73.795	ug/l	90
18) Methyl Acetate	2.709	43	4617	1.591	ug/l	94
20) Methylene Chloride	2.788	84	17118	8.052	ug/l	93
25) 2-Butanone	4.574	43	8585	5.207	ug/l	94
37) Ethyl Acetate	4.727	43	31677	9.017	ug/l	96
43) Isopropyl Acetate	6.348	43	128341	24.920	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
Data File : VX030368.D
Acq On : 03 Aug 2022 14:13
Operator : JC/MD
Sample : N3976-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP3-0729

Quant Time: Aug 03 15:55:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

