

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
 Data File : VX030369.D
 Acq On : 03 Aug 2022 14:36
 Operator : JC/MD
 Sample : N3976-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP1-0729

Quant Time: Aug 03 15:55:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	103438	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	191839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	124036	59.660	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.320%
35) Dibromofluoromethane	5.391	113	91647	49.974	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.940%
50) Toluene-d8	8.653	98	332884	46.872	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.740%
62) 4-Bromofluorobenzene	11.085	95	119487	45.334	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.660%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	76700	72.058	ug/l	90
20) Methylene Chloride	2.788	84	21462	9.721	ug/l	88
25) 2-Butanone	4.587	43	9508	5.316	ug/l	96
37) Ethyl Acetate	4.727	43	33356	8.925	ug/l	97
43) Isopropyl Acetate	6.348	43	138720	25.320	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
Data File : VX030369.D
Acq On : 03 Aug 2022 14:36
Operator : JC/MD
Sample : N3976-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP1-0729

Quant Time: Aug 03 15:55:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

