

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
 Data File : VX030371.D
 Acq On : 03 Aug 2022 15:23
 Operator : JC/MD
 Sample : N3976-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP3-0729

Quant Time: Aug 03 15:56:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	106792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	192368	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114509	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	119361	55.608	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.220%
35) Dibromofluoromethane	5.391	113	88359	48.049	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.653	98	318270	44.691	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.380%
62) 4-Bromofluorobenzene	11.086	95	115503	43.702	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.400%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	74231	67.549	ug/l	90
20) Methylene Chloride	2.788	84	17920	7.350	ug/l #	86
25) 2-Butanone	4.574	43	8914	4.827	ug/l	90
37) Ethyl Acetate	4.721	43	31160	8.315	ug/l	98
43) Isopropyl Acetate	6.349	43	131706	23.973	ug/l	96
95) Naphthalene	13.780	128	21895	2.413	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
Data File : VX030371.D
Acq On : 03 Aug 2022 15:23
Operator : JC/MD
Sample : N3976-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP3-0729

Quant Time: Aug 03 15:56:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

