

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
 Data File : VX030372.D
 Acq On : 03 Aug 2022 15:47
 Operator : JC/MD
 Sample : N3988-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 63022-C

Quant Time: Aug 04 04:41:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	79597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	158770	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	120282	75.183	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	150.360%#
35) Dibromofluoromethane	5.391	113	87258	57.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.980%
50) Toluene-d8	8.653	98	262587	44.675	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.340%
62) 4-Bromofluorobenzene	11.086	95	93806	43.003	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.000%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	7959	20.395	ug/l #	78
16) Acetone	2.386	43	90992	111.090	ug/l	89
18) Methyl Acetate	2.709	43	5792	2.391	ug/l	91
20) Methylene Chloride	2.788	84	22447	14.171	ug/l	93
25) 2-Butanone	4.574	43	12228	8.884	ug/l #	82
37) Ethyl Acetate	4.727	43	35443	11.459	ug/l	99
40) Benzene	6.050	78	122122	17.770	ug/l	97
43) Isopropyl Acetate	6.349	43	140048	30.886	ug/l	95
52) Toluene	8.720	92	123731	29.510	ug/l	99
67) Ethyl Benzene	10.195	91	45129	6.526	ug/l	94
68) m/p-Xylenes	10.305	106	88973	33.224	ug/l	88
69) o-Xylene	10.647	106	104644	39.646	ug/l	87
78) n-propylbenzene	11.305	91	7686	1.037	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	24085	4.633	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	108275	20.876	ug/l	76
95) Naphthalene	13.780	128	1739832	276.115	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
Data File : VX030372.D
Acq On : 03 Aug 2022 15:47
Operator : JC/MD
Sample : N3988-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
63022-C

Quant Time: Aug 04 04:41:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

