

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
 Data File : VX030377.D
 Acq On : 03 Aug 2022 17:44
 Operator : JC/MD
 Sample : N4018-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33773-74

Quant Time: Aug 04 04:44:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102213	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	185262	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178709	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107688	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	119793	58.310	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	116.620%	
35) Dibromofluoromethane	5.397	113	91183	51.486	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.980%	
50) Toluene-d8	8.653	98	324119	47.258	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.520%	
62) 4-Bromofluorobenzene	11.085	95	117507	46.165	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	92.340%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.946	59	10742	21.435	ug/l	# 79
16) Acetone	2.386	43	26354	25.056	ug/l	# 88
95) Naphthalene	13.780	128	9326	1.093	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
Data File : VX030377.D
Acq On : 03 Aug 2022 17:44
Operator : JC/MD
Sample : N4018-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33773-74

Quant Time: Aug 04 04:44:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

