



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080322\
 Data File : VX030360.D
 Acq On : 03 Aug 2022 10:12
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Title : SW846 8260
 Last Update : Mon Jul 25 16:49:17 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	11099	PASS
75	95	30	60	55.6	28685	PASS
95	95	100	100	100.0	51629	PASS
96	95	5	9	6.6	3410	PASS
173	174	0.00	2	1.0	339	PASS
174	95	50	100	69.0	35619	PASS
175	174	5	9	7.5	2674	PASS
176	174	95	101	95.9	34144	PASS
177	176	5	9	6.2	2131	PASS