

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080323\
 Data File : VX036794.D
 Acq On : 03 Aug 2023 12:55
 Operator : JC/MD
 Sample : 03882-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1763

Quant Time: Aug 04 07:32:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	148858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	270453	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	101063	45.884	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.760%
35) Dibromofluoromethane	5.385	113	86339	46.445	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.900%
50) Toluene-d8	8.647	98	322697	48.951	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.900%
62) 4-Bromofluorobenzene	11.079	95	121396	48.810	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.620%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3317	3.629	ug/l #	88
20) Methylene Chloride	2.794	84	2397	1.179	ug/l #	79
29) Tetrahydrofuran	5.032	42	2007	2.099	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080323\
Data File : VX036794.D
Acq On : 03 Aug 2023 12:55
Operator : JC/MD
Sample : 03882-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1763

Quant Time: Aug 04 07:32:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

