

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080323\
 Data File : VX036795.D
 Acq On : 03 Aug 2023 13:19
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 04 07:32:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	262743	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108391	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99254	46.800	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.600%
35) Dibromofluoromethane	5.385	113	83877	46.445	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.880%
50) Toluene-d8	8.647	98	313285	48.918	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.079	95	117909	48.799	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.600%
Target Compounds						
20) Methylene Chloride	2.788	84	1371	0.700	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080323\
Data File : VX036795.D
Acq On : 03 Aug 2023 13:19
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 04 07:32:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

