

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080323\
 Data File : VX036796.D
 Acq On : 03 Aug 2023 13:43
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda

Quant Time: Aug 04 07:32:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	146529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	253800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110370	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104935	48.399	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.800%
35) Dibromofluoromethane	5.385	113	86981	49.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.720%
50) Toluene-d8	8.647	98	298624	48.272	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.540%
62) 4-Bromofluorobenzene	11.079	95	115456	49.467	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.940%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	75432	38.969	ug/l	100
3) Chloromethane	1.300	50	69115	39.257	ug/l	99
4) Vinyl Chloride	1.374	62	77218	43.078	ug/l	100
5) Bromomethane	1.599	94	55539	51.898	ug/l	96
6) Chloroethane	1.678	64	43760	50.213	ug/l	100
7) Trichlorofluoromethane	1.880	101	140285	50.679	ug/l	100
8) Diethyl Ether	2.136	74	51843	48.409	ug/l	77
9) 1,1,2-Trichlorotrifluo...	2.325	101	77639	46.684	ug/l	98
10) Methyl Iodide	2.453	142	84998	49.207	ug/l	93
11) Tert butyl alcohol	2.965	59	130751	254.732	ug/l #	92
12) 1,1-Dichloroethene	2.312	96	74674	45.512	ug/l	88
13) Acrolein	2.239	56	77563	264.781	ug/l	99
14) Allyl chloride	2.660	41	134067	41.495	ug/l #	96
15) Acrylonitrile	3.062	53	251587	243.728	ug/l	99
16) Acetone	2.380	43	216298	240.415	ug/l	92
17) Carbon Disulfide	2.508	76	167288	36.942	ug/l	99
18) Methyl Acetate	2.703	43	183728	48.544	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	305743	49.587	ug/l	98
20) Methylene Chloride	2.788	84	88910	44.436	ug/l	92
21) trans-1,2-Dichloroethene	3.093	96	82456	44.965	ug/l	91
22) Diisopropyl ether	3.763	45	279121	45.500	ug/l #	83
23) Vinyl Acetate	3.721	43	1077016	228.877	ug/l #	95
24) 1,1-Dichloroethane	3.611	63	162981	47.386	ug/l	98
25) 2-Butanone	4.556	43	355294	243.404	ug/l	93
26) 2,2-Dichloropropane	4.477	77	136987	42.269	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	103997	47.748	ug/l	92
28) Bromochloromethane	4.897	49	77690	46.705	ug/l	87
29) Tetrahydrofuran	5.007	42	224194	238.229	ug/l	89
30) Chloroform	5.092	83	175556	49.891	ug/l	96
31) Cyclohexane	5.470	56	123553	41.158	ug/l	92
32) 1,1,1-Trichloroethane	5.385	97	156549	48.554	ug/l	98
36) 1,1-Dichloropropene	5.690	75	118683	45.645	ug/l	98
37) Ethyl Acetate	4.714	43	138536	47.074	ug/l	97
38) Carbon Tetrachloride	5.678	117	134867	48.789	ug/l	98
39) Methylcyclohexane	7.379	83	137398	42.496	ug/l	95
40) Benzene	6.037	78	353790	46.767	ug/l	100

08/04/2023
 Supervised By :Semsettin
 Esilyurt

08/04/2023

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080323\
 Data File : VX036796.D
 Acq On : 03 Aug 2023 13:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda

Quant Time: Aug 04 07:32:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
41) Methacrylonitrile	4.922	41	73950	46.982	ug/l	#	91
42) 1,2-Dichloroethane	6.086	62	138760	50.169	ug/l		96
43) Isopropyl Acetate	6.342	43	231762	45.749	ug/l	#	95
44) Trichloroethene	7.123	130	96282	48.691	ug/l		99
45) 1,2-Dichloropropane	7.427	63	94787	46.930	ug/l		98
46) Dibromomethane	7.580	93	69006	48.817	ug/l		95
47) Bromodichloromethane	7.824	83	140356	48.372	ug/l		98
48) Methyl methacrylate	7.690	41	109477	46.469	ug/l		89
49) 1,4-Dioxane	7.659	88	49740	1045.093	ug/l	#	90
51) 4-Methyl-2-Pentanone	8.574	43	682037	243.285	ug/l		95
52) Toluene	8.720	92	228138	48.129	ug/l		99
53) t-1,3-Dichloropropene	8.976	75	152224	46.417	ug/l		100
54) cis-1,3-Dichloropropene	8.366	75	159405	46.063	ug/l		91
55) 1,1,2-Trichloroethane	9.153	97	98654	49.512	ug/l		96
56) Ethyl methacrylate	9.116	69	157485	48.538	ug/l		94
57) 1,3-Dichloropropane	9.311	76	164977	49.711	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.244	63	375911	235.135	ug/l		96
59) 2-Hexanone	9.433	43	512092	246.125	ug/l		93
60) Dibromochloromethane	9.518	129	110063	50.256	ug/l		100
61) 1,2-Dibromoethane	9.610	107	104917	50.805	ug/l		100
64) Tetrachloroethene	9.275	164	81795	48.645	ug/l		99
65) Chlorobenzene	10.079	112	246336	49.573	ug/l		100
66) 1,1,1,2-Tetrachloroethane	10.165	131	98331	51.197	ug/l		99
67) Ethyl Benzene	10.195	91	439136	48.375	ug/l		99
68) m/p-Xylenes	10.299	106	330991	97.724	ug/l		99
69) o-Xylene	10.640	106	166740	49.320	ug/l		99
70) Styrene	10.652	104	276340	49.542	ug/l		98
71) Bromoform	10.799	173	78685	50.610	ug/l	#	99
73) Isopropylbenzene	10.963	105	441062	48.287	ug/l		99
74) N-amyl acetate	10.841	43	190414	44.704	ug/l		96
75) 1,1,2,2-Tetrachloroethane	11.213	83	149653	48.509	ug/l		100
76) 1,2,3-Trichloropropane	11.238	75	126474m	48.718	ug/l		
77) Bromobenzene	11.195	156	102607	49.327	ug/l		95
78) n-propylbenzene	11.305	91	500158	47.347	ug/l		99
79) 2-Chlorotoluene	11.366	91	302507	47.258	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	369565	48.364	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	49590	43.296	ug/l		87
82) 4-Chlorotoluene	11.451	91	348543	47.788	ug/l		99
83) tert-Butylbenzene	11.713	119	365190	48.648	ug/l		99
84) 1,2,4-Trimethylbenzene	11.750	105	367651	48.109	ug/l		97
85) sec-Butylbenzene	11.890	105	460577	49.198	ug/l		99
86) p-Isopropyltoluene	12.006	119	383237	49.750	ug/l		98
87) 1,3-Dichlorobenzene	11.969	146	190608	49.474	ug/l		99
88) 1,4-Dichlorobenzene	12.042	146	191944	50.202	ug/l		99
89) n-Butylbenzene	12.329	91	332953	48.283	ug/l		99
90) Hexachloroethane	12.536	117	72566	46.467	ug/l		97
91) 1,2-Dichlorobenzene	12.335	146	188836	50.792	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	35535	49.568	ug/l		86
93) 1,2,4-Trichlorobenzene	13.585	180	117689	51.380	ug/l		97
94) Hexachlorobutadiene	13.725	225	45602	48.123	ug/l		97
95) Naphthalene	13.774	128	414119	51.881	ug/l		100
96) 1,2,3-Trichlorobenzene	13.963	180	118414	52.452	ug/l		99

08/04/2023
 Supervised By :Semsettin
 Yesilyurt

08/04/2023

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080323\
Data File : VX036796.D
Acq On : 03 Aug 2023 13:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Quant Time: Aug 04 07:32:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						


08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 732
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 146529
Ion Ratio Lower Upper
168 100
99 62.1 56.6 85.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#2
Dichlorodifluoromethane
Concen: 38.969 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion: 85 Resp: 75432
Ion Ratio Lower Upper
85 100
87 33.6 16.7 50.1





#3
Chloromethane
Concen: 39.257 ug/l
RT: 1.300 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 69115
Ion Ratio Lower Upper
50 100
52 31.3 25.5 38.3


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#4
Vinyl Chloride
Concen: 43.078 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion: 62 Resp: 77218
Ion Ratio Lower Upper
62 100
64 32.0 25.4 38.0





#5
Bromomethane
Concen: 51.898 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.018 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 55539
Ion Ratio Lower Upper
94 100
96 96.9 74.6 111.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#6
Chloroethane
Concen: 50.213 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.012 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion: 64 Resp: 43760
Ion Ratio Lower Upper
64 100
66 31.1 25.1 37.7





#7
Trichlorofluoromethane
Concen: 50.679 ug/l
RT: 1.880 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 14028
Ion Ratio Lower Upper
101 100
103 64.2 51.4 77.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#8
Diethyl Ether
Concen: 48.409 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion: 74 Resp: 51843
Ion Ratio Lower Upper
74 100
45 83.0 53.3 159.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.684 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 77639
Ion Ratio Lower Upper
101 100
85 43.9 37.0 55.6
151 74.0 58.6 87.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#10
Methyl Iodide
Concen: 49.207 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion:142 Resp: 84998
Ion Ratio Lower Upper
142 100
127 46.1 41.6 62.4
141 13.7 11.6 17.4





#11

Tert butyl alcohol

Concen: 254.732 ug/l

RT: 2.965 min Scan# 308

Delta R.T. -0.024 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 13075

Ion Ratio Lower Upper



59 100

57 10.2 6.0 9.0

Manual Integrations



APPROVED

Reviewed By :Mahesh
Dadoda



08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#12

1,1-Dichloroethene

Concen: 45.512 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.006 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43




Tgt Ion: 96 Resp: 74674

Ion Ratio Lower Upper

96 100

61 159.3 144.6 216.8

98 62.0 50.4 75.6







#13
Acrolein
Concen: 264.781 ug/l
RT: 2.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 56 Resp: 7756
Ion Ratio Lower Upper
56 100
55 70.4 57.1 85.7



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#14
Allyl chloride
Concen: 41.495 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion: 41 Resp: 134067
Ion Ratio Lower Upper
41 100
39 68.6 55.8 83.6
76 36.7 24.3 36.5#





#15
Acrylonitrile
Concen: 243.728 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 25158
Ion Ratio Lower Upper
53 100
52 82.6 66.0 99.0
51 35.9 29.8 44.6


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#16
Acetone
Concen: 240.415 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.012 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion: 43 Resp: 216298
Ion Ratio Lower Upper
43 100
58 31.6 21.9 32.9





#17

Carbon Disulfide

Concen: 36.942 ug/l

RT: 2.508 min Scan# 21

Delta R.T. -0.006 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 16728

Ion Ratio Lower Upper



76 100

78 9.1 7.0 10.6

Manual Integrations



APPROVED

Reviewed By :Mahesh
Dadoda



08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#18

Methyl Acetate

Concen: 48.544 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.006 min

Lab File: VX036796.D




Acq: 03 Aug 2023 13:43


Tgt Ion: 43 Resp: 183728

Ion Ratio Lower Upper

43 100

74 25.8 17.8 26.8





#19

Methyl tert-butyl Ether

Concen: 49.587 ug/l

RT: 3.117 min Scan# 311

Delta R.T. 0.006 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 73 Resp: 30574

Ion Ratio Lower Upper



73 100

57 20.9 17.3 25.9

Manual Integrations

APPROVED



Reviewed By :Mahesh
Dadoda



08/04/2023

Supervised By :Semsettin
Yesilyurt

08/04/2023



#20

Methylene Chloride

Concen: 44.436 ug/l

RT: 2.788 min Scan# 279

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Tgt Ion: 84 Resp: 88910




Ion Ratio Lower Upper


84 100

49 117.7 103.6 155.4

51 36.6 33.8 50.6


86 63.5 49.4 74.0





#21
trans-1,2-Dichloroethene
Concen: 44.965 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 82450
Ion Ratio Lower Upper
96 100
61 133.8 119.2 178.8
98 63.0 48.7 73.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#22
Diisopropyl ether
Concen: 45.500 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion: 45 Resp: 279121
Ion Ratio Lower Upper
45 100
43 67.5 69.7 104.5#
87 30.8 20.8 31.2
59 12.1 9.0 13.4





#23
 Vinyl Acetate
 Concen: 228.877 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX036796.D
 Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 1077010
 Ion Ratio Lower Upper
 43 100
 86 11.8 7.8 11.8



Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda






08/04/2023
 Supervised By :Semsettin
 Yesilyurt


08/04/2023



#24
 1,1-Dichloroethane
 Concen: 47.386 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.000 min
 Lab File: VX036796.D
 Acq: 03 Aug 2023 13:43

Tgt Ion: 63 Resp: 162981
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.3 9.8
 100 5.1 2.1 6.5







#25
2-Butanone
Concen: 243.404 ug/l
RT: 4.556 min Scan# 50
Delta R.T. -0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 355294
Ion Ratio Lower Upper
43 100
72 27.8 19.4 29.2



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#26
2,2-Dichloropropane
Concen: 42.269 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion: 77 Resp: 136987
Ion Ratio Lower Upper
77 100
97 22.6 10.8 32.3






#27
 cis-1,2-Dichloroethene
 Concen: 47.748 ug/l
 RT: 4.489 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: VX036796.D
 Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :



VSTDCCC050EC



Tgt Ion: 96 Resp: 10399
 Ion Ratio Lower Upper
 96 100
 61 142.1 0.0 312.2
 98 64.1 0.0 129.0


Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda




08/04/2023
 Supervised By :Semsettin
 Yesilyurt



08/04/2023




#28
 Bromochloromethane
 Concen: 46.705 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.006 min
 Lab File: VX036796.D
 Acq: 03 Aug 2023 13:43



Tgt Ion: 49 Resp: 77690
 Ion Ratio Lower Upper
 49 100
 129 2.2 0.0 3.8
 130 83.0 57.5 86.3





#29

Tetrahydrofuran

Concen: 238.229 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 42 Resp: 224194

Ion	Ratio	Lower	Upper
42	100		
72	50.1	34.2	51.4
71	46.1	32.2	48.2

Manual Integrations



APPROVED

Reviewed By :Mahesh
Dadoda



08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#30

Chloroform

Concen: 49.891 ug/l

RT: 5.092 min Scan# 657




Delta R.T. -0.006 min


Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Tgt Ion: 83 Resp: 175556


Ion	Ratio	Lower	Upper
83	100		
85	65.7	50.3	75.5





#31
Cyclohexane
Concen: 41.158 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 12355
Ion Ratio Lower Upper
56 100
69 32.3 23.4 35.2
84 89.8 65.9 98.9


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#32
1,1,1-Trichloroethane
Concen: 48.554 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion: 97 Resp: 156549
Ion Ratio Lower Upper
97 100
99 64.5 51.1 76.7
61 43.0 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 48.399 ug/l

RT: 5.958 min Scan# 798

Delta R.T. 0.006 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

Client SampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 104935

Ion Ratio Lower Upper



65 100

67 51.6 0.0 100.2

Manual Integrations

APPROVED



Reviewed By :Mahesh
Dadoda



08/04/2023

Supervised By :Semsettin
Yesilyurt

08/04/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43




Tgt Ion:114 Resp: 253800


Ion Ratio Lower Upper

114 100

63 20.3 0.0 45.0


88 16.3 0.0 36.6





#35
Dibromofluoromethane
Concen: 49.861 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 113 Resp: 8698
Ion Ratio Lower Upper
113 100
111 103.3 82.9 124.3
192 18.1 13.3 19.9


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#36
1,1-Dichloropropene
Concen: 45.645 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion: 75 Resp: 118683
Ion Ratio Lower Upper
75 100
110 35.9 17.0 50.9
77 31.0 24.7 37.1





#37
Ethyl Acetate
Concen: 47.074 ug/l
RT: 4.714 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 138530
Ion Ratio Lower Upper
43 100
61 14.6 10.6 16.0
70 11.7 8.3 12.5



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#38
Carbon Tetrachloride
Concen: 48.789 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion:117 Resp: 134867
Ion Ratio Lower Upper
117 100
119 95.2 78.2 117.2
121 31.7 24.4 36.6





#39

Methylcyclohexane

Concen: 42.496 ug/l

RT: 7.379 min Scan# 1033

Delta R.T. -0.006 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 83 Resp: 137398

Ion	Ratio	Lower	Upper
83	100		
55	75.9	64.6	97.0
98	49.3	36.7	55.1

Manual Integrations



APPROVED

Reviewed By :Mahesh
Dadoda



08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#40

Benzene

Concen: 46.767 ug/l

RT: 6.037 min Scan# 812




Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Tgt Ion: 78 Resp: 353790

Ion	Ratio	Lower	Upper
78	100		
77	24.2	19.3	28.9






#41
Methacrylonitrile
Concen: 46.982 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :



VSTDCCC050EC



Tgt Ion:	41	Resp:	73950
Ion Ratio	100	Lower	Upper
39	57.9	45.2	67.8
67	78.9	52.6	78.8
52	31.2	22.7	34.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#42
1,2-Dichloroethane
Concen: 50.169 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion:	62	Resp:	138760
Ion Ratio	100	Lower	Upper
98	9.3	0.0	16.0





#43
Isopropyl Acetate
Concen: 45.749 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
Client SampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 231762
Ion Ratio Lower Upper
43 100
61 20.9 15.2 22.8
87 14.7 9.8 14.6


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#44
Trichloroethene
Concen: 48.691 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion:130 Resp: 96282
Ion Ratio Lower Upper
130 100
95 100.9 0.0 204.4





#45

1,2-Dichloropropane

Concen: 46.930 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 9478

Ion Ratio Lower Upper



63 100

65 29.4 24.3 36.5

Manual Integrations

APPROVED



Reviewed By :Mahesh
Dadoda



08/04/2023

Supervised By :Semsettin
Yesilyurt

08/04/2023



#46

Dibromomethane

Concen: 48.817 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43




Tgt Ion: 93 Resp: 69006


Ion Ratio Lower Upper

93 100

95 84.7 66.6 100.0

174 93.9 69.4 104.0





#47
Bromodichloromethane
Concen: 48.372 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 140350
Ion Ratio Lower Upper
83 100
85 64.1 49.8 74.6
127 8.9 7.1 10.7



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#48
Methyl methacrylate
Concen: 46.469 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion: 41 Resp: 109477
Ion Ratio Lower Upper
41 100
69 87.7 59.2 88.8
39 62.7 47.7 71.5





#49

1,4-Dioxane

Concen: 1045.093 ug/l

RT: 7.659 min Scan# 1084

Delta R.T. -0.037 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 88 Resp: 49740

Ion	Ratio	Lower	Upper
88	100		
43	32.1	0.0	0.0
58	66.5	60.1	90.1

Manual Integrations



APPROVED

Reviewed By :Mahesh
Dadoda



08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#50

Toluene-d8

Concen: 48.272 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. -0.000 min


Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Tgt Ion: 98 Resp: 298624

Ion	Ratio	Lower	Upper
98	100		
100	65.7	53.1	79.7







#51
4-Methyl-2-Pentanone
Concen: 243.285 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 68203
Ion Ratio Lower Upper
43 100
58 41.3 30.4 45.6



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#52
Toluene
Concen: 48.129 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion: 92 Resp: 228138
Ion Ratio Lower Upper
92 100
91 169.4 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 46.417 ug/l

RT: 8.976 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 75 Resp: 152224

Ion Ratio Lower Upper



75 100

77 32.3 25.7 38.5

Manual Integrations

APPROVED



Reviewed By :Mahesh
Dadoda



08/04/2023

Supervised By :Semsettin
Yesilyurt

08/04/2023



#54

cis-1,3-Dichloropropene

Concen: 46.063 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43




Tgt Ion: 75 Resp: 159405


Ion Ratio Lower Upper

75 100

77 32.6 25.3 37.9

39 48.5 46.7 70.1







#55
1,1,2-Trichloroethane
Concen: 49.512 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:	97	Resp:	98654
Ion	Ratio	Lower	Upper
97	100		
83	85.5	65.8	98.8
85	54.8	41.6	62.4
99	62.9	47.8	71.8



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#56
Ethyl methacrylate
Concen: 48.538 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion:	69	Resp:	157485
Ion	Ratio	Lower	Upper
69	100		
41	63.2	56.8	85.2
39	39.2	32.0	48.0





#57

1,3-Dichloropropane

Concen: 49.711 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.006 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 16497

Ion Ratio Lower Upper



76 100

78 32.0 26.1 39.1

Manual Integrations

APPROVED



Reviewed By :Mahesh
Dadoda



08/04/2023

Supervised By :Semsettin
Yesilyurt

08/04/2023



#58

2-Chloroethyl Vinyl ether

Concen: 235.135 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX036796.D




Acq: 03 Aug 2023 13:43

Tgt Ion: 63 Resp: 375911

Ion Ratio Lower Upper

63 100

106 30.1 22.5 33.7







#59
2-Hexanone
Concen: 246.125 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 512092
Ion Ratio Lower Upper
43 100
58 57.3 26.2 78.5



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#60
Dibromochloromethane
Concen: 50.256 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion:129 Resp: 110063
Ion Ratio Lower Upper
129 100
127 77.4 38.8 116.4





#61
1,2-Dibromoethane
Concen: 50.805 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:107 Resp: 104917
Ion Ratio Lower Upper
107 100
109 94.5 75.2 112.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#62
4-Bromofluorobenzene
Concen: 49.467 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43




Tgt Ion: 95 Resp: 115456
Ion Ratio Lower Upper
95 100
174 73.7 0.0 135.6
176 72.4 0.0 131.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 225432
Ion Ratio Lower Upper
117 100
82 55.6 46.2 69.4
119 32.1 25.5 38.3


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#64
Tetrachloroethene
Concen: 48.645 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43




Tgt Ion:164 Resp: 81795
Ion Ratio Lower Upper
164 100
166 128.0 102.7 154.1
129 97.3 80.4 120.6
131 95.3 76.5 114.7





#65
Chlorobenzene
Concen: 49.573 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:112 Resp: 246330
Ion Ratio Lower Upper
112 100
114 31.9 25.7 38.5


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#66
1,1,1,2-Tetrachloroethane
Concen: 51.197 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43




Tgt Ion:131 Resp: 98331
Ion Ratio Lower Upper
131 100
133 94.8 47.1 141.4
119 66.1 32.2 96.6





#67
Ethyl Benzene
Concen: 48.375 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 439136
Ion Ratio Lower Upper
91 100
106 31.4 24.9 37.3


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#68
m/p-Xylenes
Concen: 97.724 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion:106 Resp: 330991
Ion Ratio Lower Upper
106 100
91 205.8 165.5 248.3







#69
o-Xylene
Concen: 49.320 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 166740
Ion Ratio Lower Upper
106 100
91 215.9 108.9 326.8



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#70
Styrene
Concen: 49.542 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Tgt Ion:104 Resp: 276340
Ion Ratio Lower Upper
104 100
78 52.8 43.6 65.4
103 55.2 43.6 65.4





#71

Bromoform

Concen: 50.610 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :



VSTDCCC050EC

Tgt Ion:173	Resp:	7868
Ion Ratio	Lower	Upper
173	100	
175	47.4	24.1 72.4
254	0.1	0.0 0.0

Manual Integrations



APPROVED

Reviewed By :Mahesh
Dadoda



08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Tgt Ion:152	Resp:	110370
Ion Ratio	Lower	Upper
152	100	
115	88.3	43.3 129.9
150	173.4	0.0 345.0





#73
Isopropylbenzene
Concen: 48.287 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 441062
Ion Ratio Lower Upper
105 100
120 26.4 12.9 38.6



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#74
N-ethyl acetate
Concen: 44.704 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion: 43 Resp: 190414
Ion Ratio Lower Upper
43 100
70 45.2 32.9 49.3
55 24.1 19.3 28.9
61 25.4 18.8 28.2





#75
1,1,2,2-Tetrachloroethane
Concen: 48.509 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 14965
Ion Ratio Lower Upper
83 100
131 10.1 5.3 15.8
85 64.4 32.0 96.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#76
1,2,3-Trichloropropane
Concen: 48.718 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43




Tgt Ion: 75 Resp: 126474
Ion Ratio Lower Upper
75 100
77 43.4 19.1 57.5





#77
Bromobenzene
Concen: 49.327 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:156 Resp: 102607
Ion Ratio Lower Upper
156 100
77 166.3 88.4 265.3
158 95.1 48.0 144.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023



#78
n-propylbenzene
Concen: 47.347 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion: 91 Resp: 500158
Ion Ratio Lower Upper
91 100
120 23.3 11.5 34.4







#79
2-Chlorotoluene
Concen: 47.258 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 30250
Ion Ratio Lower Upper
91 100
126 33.3 16.3 48.8



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#80
1,3,5-Trimethylbenzene
Concen: 48.364 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion:105 Resp: 369565
Ion Ratio Lower Upper
105 100
120 48.4 24.1 72.3





#81
trans-1,4-Dichloro-2-butene
Concen: 43.296 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 49590
Ion Ratio Lower Upper
75 100
53 94.2 90.5 135.7
89 45.2 36.9 55.3


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#82
4-Chlorotoluene
Concen: 47.788 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion: 91 Resp: 348543
Ion Ratio Lower Upper
91 100
126 28.9 14.2 42.6





#83
tert-Butylbenzene
Concen: 48.648 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:119 Resp: 365190
Ion Ratio Lower Upper
119 100
91 61.2 31.2 93.6
134 23.2 11.7 35.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#84
1,2,4-Trimethylbenzene
Concen: 48.109 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion:105 Resp: 367651
Ion Ratio Lower Upper
105 100
120 44.6 21.2 63.6





#85
 sec-Butylbenzene
 Concen: 49.198 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX036796.D
 Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 46057
 Ion Ratio Lower Upper
 105 100
 134 20.0 9.8 29.4



Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda






08/04/2023
 Supervised By :Semsettin
 Yesilyurt

08/04/2023



#86
 p-Isopropyltoluene
 Concen: 49.750 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX036796.D
 Acq: 03 Aug 2023 13:43

Tgt Ion:119 Resp: 383237
 Ion Ratio Lower Upper
 119 100
 134 26.2 12.8 38.3
 91 24.9 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 49.474 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 190608

Ion Ratio Lower Upper

146 100



111 42.4 22.1 66.1

148 63.7 31.8 95.3

Manual Integrations

APPROVED



Reviewed By :Mahesh
Dadoda



08/04/2023

Supervised By :Semsettin
Yesilyurt

08/04/2023



#88

1,4-Dichlorobenzene

Concen: 50.202 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43




Tgt Ion:146 Resp: 191944


Ion Ratio Lower Upper

146 100

111 41.3 21.3 63.7


148 63.2 31.8 95.4





#89
n-Butylbenzene
Concen: 48.283 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 33295
Ion Ratio Lower Upper
91 100
92 54.1 27.4 82.0
134 25.9 12.6 37.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






08/04/2023
Supervised By :Semsettin
Yesilyurt


08/04/2023



#90
Hexachloroethane
Concen: 46.467 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43


Tgt Ion:117 Resp: 72566
Ion Ratio Lower Upper
117 100
201 68.2 32.9 98.6





#91
1,2-Dichlorobenzene
Concen: 50.792 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 188830
Ion Ratio Lower Upper
146 100
111 43.2 22.4 67.0
148 63.2 31.6 95.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt



08/04/2023




#92
1,2-Dibromo-3-Chloropropane
Concen: 49.568 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion: 75 Resp: 35535
Ion Ratio Lower Upper
75 100
155 82.7 36.1 108.5
157 104.8 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 51.380 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion:180	Resp:	117689
Ion Ratio	Lower	Upper
180	100	
182	94.2	48.3 144.8
145	33.1	18.0 54.0



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#94

Hexachlorobutadiene

Concen: 48.123 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. -0.000 min

Lab File: VX036796.D

Acq: 03 Aug 2023 13:43


Tgt Ion:225	Resp:	45602
Ion Ratio	Lower	Upper
225	100	
223	62.5	31.1 93.2
227	61.0	32.5 97.5





#95
Naphthalene
Concen: 51.881 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 414119
Ion Ratio Lower Upper
128 100
127 13.0 10.3 15.5
129 11.2 8.8 13.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




08/04/2023
Supervised By :Semsettin
Yesilyurt

08/04/2023



#96
1,2,3-Trichlorobenzene
Concen: 52.452 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX036796.D
Acq: 03 Aug 2023 13:43



Tgt Ion:180 Resp: 118414
Ion Ratio Lower Upper
180 100
182 94.0 46.9 140.6
145 34.2 17.8 53.5

