

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX080420\
 Data File : VX017771.D
 Acq On : 04 Aug 2020 15:25
 Operator : JC/SP
 Sample : L3556-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB179-GW-398-400

Quant Time: Aug 05 02:58:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	582255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	891537	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	821097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	411711	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	375583	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
35) Dibromofluoromethane	5.47	113	293034	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
50) Toluene-d8	8.70	98	1042886	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
62) 4-Bromofluorobenzene	11.12	95	389409	45.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
12) 1,1-Dichloroethene	2.36	96	5426	0.902	ug/l	90
16) Acetone	2.42	43	14401	4.375	ug/l #	78
21) trans-1,2-Dichloroethene	3.15	96	3736	0.580	ug/l	87
24) 1,1-Dichloroethane	3.68	63	6180	0.534	ug/l #	90
27) cis-1,2-Dichloroethene	4.57	96	203793	29.122	ug/l	86
42) 1,2-Dichloroethane	6.17	62	2764	0.276	ug/l #	74
44) Trichloroethene	7.19	130	1830851	255.559	ug/l	98
45) 1,2-Dichloropropane	7.49	63	11594	1.706	ug/l #	88
55) 1,1,2-Trichloroethane	9.20	97	1869	0.285	ug/l #	91
64) Tetrachloroethene	9.32	164	131799	17.884	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX080420\
Data File : VX017771.D
Acq On : 04 Aug 2020 15:25
Operator : JC/SP
Sample : L3556-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB179-GW-398-400

Quant Time: Aug 05 02:58:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

