

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX080420\
 Data File : VX017776.D
 Acq On : 04 Aug 2020 17:19
 Operator : JC/SP
 Sample : PB130720ZHE#17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130720ZHE#17

Quant Time: Aug 05 03:12:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	593868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	891078	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	829038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	408102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	390277	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
35) Dibromofluoromethane	5.47	113	302324	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
50) Toluene-d8	8.70	98	1049371	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	11.12	95	397942	46.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.50%	
Target Compounds						
16) Acetone	2.42	43	39225	11.683	ug/l	98
18) Methyl Acetate	2.75	43	9109	1.152	ug/l #	79
20) Methylene Chloride	2.84	84	34043	4.207	ug/l	95
43) Isopropyl Acetate	6.42	43	155620	10.087	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX080420\
Data File : VX017776.D
Acq On : 04 Aug 2020 17:19
Operator : JC/SP
Sample : PB130720ZHE#17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130720ZHE#17

Quant Time: Aug 05 03:12:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

