

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX080420\
 Data File : VX017779.D
 Acq On : 04 Aug 2020 18:28
 Operator : JC/SP
 Sample : PB130720ZHE#20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130720ZHE#20

Quant Time: Aug 05 03:37:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	592167	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	895563	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	852625	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	403463	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	389083	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
35) Dibromofluoromethane	5.47	113	303599	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	8.70	98	1106888	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
62) 4-Bromofluorobenzene	11.12	95	409498	47.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.72%	
Target Compounds						
16) Acetone	2.42	43	37020	11.058	ug/l	98
18) Methyl Acetate	2.75	43	10530	1.335	ug/l #	78
20) Methylene Chloride	2.84	84	30847	3.703	ug/l	94
43) Isopropyl Acetate	6.42	43	163228	10.527	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX080420\
Data File : VX017779.D
Acq On : 04 Aug 2020 18:28
Operator : JC/SP
Sample : PB130720ZHE#20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130720ZHE#20

Quant Time: Aug 05 03:37:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

