



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080421\
 Data File : VX023531.D
 Acq On : 04 Aug 2021 08:55
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Tue Aug 03 12:05:42 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	11652	PASS
75	95	30	60	53.1	31189	PASS
95	95	100	100	100.0	58781	PASS
96	95	5	9	6.9	4043	PASS
173	174	0.00	2	0.8	423	PASS
174	95	50	100	87.1	51173	PASS
175	174	5	9	7.1	3635	PASS
176	174	95	101	95.0	48611	PASS
177	176	5	9	6.3	3075	PASS