

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080422\
 Data File : VX030410.D
 Acq On : 04 Aug 2022 22:31
 Operator : JC/MD
 Sample : N4021-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 F5E16

Quant Time: Aug 05 05:30:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 05 05:25:14 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.769	114	253843	50.000	ug/L	0.00
28) Chlorobenzene-d5		10.055	117	226283	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		12.024	152	100597	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.368	65	96228	59.031	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	118.060%	
7) Chloroethane-d5		1.666	69	76008	66.142	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	132.280%#	
11) 1,1-Dichloroethene-d2		2.307	63	153688	49.841	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.680%	
21) 2-Butanone-d5		4.471	46	190657	126.229	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	126.230%	
24) Chloroform-d		5.062	84	195914	58.383	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	116.760%	
26) 1,2-Dichloroethane-d4		5.964	65	139434	61.195	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	122.380%	
32) Benzene-d6		5.977	84	398263	61.443	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	122.880%	
36) 1,2-Dichloropropane-d6		7.312	67	127480	60.619	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	121.240%#	
41) Toluene-d8		8.653	98	358258	58.636	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	117.280%	
43) trans-1,3-Dichloroprop...		8.952	79	56829	54.320	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.640%	
47) 2-Hexanone-d5		9.384	63	132282	123.405	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	123.410%	
56) 1,1,2,2-Tetrachloroeth...		11.195	84	167067	60.474	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	120.940%#	
66) 1,2-Dichlorobenzene-d4		12.323	152	120682	61.851	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	123.700%#	
Target Compounds							
5) Vinyl chloride		1.374	62	83652	47.682	ug/L	98
17) trans-1,2-Dichloroethene		3.093	96	3675	2.233	ug/L	89
20) cis-1,2-Dichloroethene		4.495	96	55656	30.058	ug/L	98
34) Trichloroethene		7.129	95	21622	10.290	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080422\
Data File : VX030410.D
Acq On : 04 Aug 2022 22:31
Operator : JC/MD
Sample : N4021-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
F5E16

Quant Time: Aug 05 05:30:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 05 05:25:14 2022
Response via : Initial Calibration

