

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
 Data File : VX036800.D
 Acq On : 04 Aug 2023 11:03
 Operator : JC/MD
 Sample : VX0804WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0804WBL01

Quant Time: Aug 04 23:05:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	137537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	247424	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98336	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95077	46.719	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.440%
35) Dibromofluoromethane	5.385	113	80457	47.309	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.620%
50) Toluene-d8	8.647	98	295063	48.925	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.079	95	110041	48.362	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
Data File : VX036800.D
Acq On : 04 Aug 2023 11:03
Operator : JC/MD
Sample : VX0804WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0804WBL01

Quant Time: Aug 04 23:05:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

