

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
 Data File : VX036810.D
 Acq On : 04 Aug 2023 15:07
 Operator : JC/MD
 Sample : ZHC#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ZHC#08

Quant Time: Aug 05 00:12:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	137822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98054	48.083	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.160%		
35) Dibromofluoromethane	5.385	113	81526	47.395	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.780%		
50) Toluene-d8	8.647	98	303310	49.723	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.440%		
62) 4-Bromofluorobenzene	11.079	95	114820	49.890	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	99.780%		
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	41582	49.138	ug/l	91
20) Methylene Chloride	2.794	84	7382	3.922	ug/l #	87
25) 2-Butanone	4.568	43	18542	13.505	ug/l	98
37) Ethyl Acetate	4.721	43	6286	2.166	ug/l #	93
43) Isopropyl Acetate	6.342	43	135487	27.123	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
Data File : VX036810.D
Acq On : 04 Aug 2023 15:07
Operator : JC/MD
Sample : ZHC#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ZHC#08

Quant Time: Aug 05 00:12:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

