

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
 Data File : VX036813.D
 Acq On : 04 Aug 2023 16:19
 Operator : JC/MD
 Sample : ZHC#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ZHC#11

Quant Time: Aug 05 00:13:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	157439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	287532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	269406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	121694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108753	46.684	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.360%
35) Dibromofluoromethane	5.385	113	92648	46.879	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.760%
50) Toluene-d8	8.647	98	345693	49.325	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.640%
62) 4-Bromofluorobenzene	11.079	95	132135	49.972	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.940%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	37841	39.146	ug/l	92
20) Methylene Chloride	2.788	84	6324	2.942	ug/l	95
25) 2-Butanone	4.562	43	17229	10.985	ug/l	94
37) Ethyl Acetate	4.721	43	5424	1.627	ug/l #	83
43) Isopropyl Acetate	6.342	43	125901	21.937	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
Data File : VX036813.D
Acq On : 04 Aug 2023 16:19
Operator : JC/MD
Sample : ZHC#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ZHC#11

Quant Time: Aug 05 00:13:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

