

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
 Data File : VX036817.D
 Acq On : 04 Aug 2023 17:55
 Operator : JC/MD
 Sample : 03877-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-2

Quant Time: Aug 04 23:09:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	244026	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95962	48.257	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.520%
35) Dibromofluoromethane	5.385	113	80137	47.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.560%
50) Toluene-d8	8.647	98	296046	49.772	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.079	95	110597	49.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.560%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	40200	48.717	ug/l	93
20) Methylene Chloride	2.788	84	7230	3.940	ug/l	84
25) 2-Butanone	4.562	43	17564	13.119	ug/l #	83
37) Ethyl Acetate	4.721	43	6042	2.135	ug/l #	94
43) Isopropyl Acetate	6.342	43	135772	27.875	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
Data File : VX036817.D
Acq On : 04 Aug 2023 17:55
Operator : JC/MD
Sample : 03877-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-2

Quant Time: Aug 04 23:09:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

