

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
 Data File : VX036820.D
 Acq On : 04 Aug 2023 19:07
 Operator : JC/MD
 Sample : 03887-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MCLEAN-TP7

Quant Time: Aug 04 23:10:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	252540	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96306	47.123	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.240%
35) Dibromofluoromethane	5.385	113	81737	47.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.180%
50) Toluene-d8	8.647	98	301338	48.953	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.900%
62) 4-Bromofluorobenzene	11.079	95	113397	48.827	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.660%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	32738	38.603	ug/l #	89
20) Methylene Chloride	2.788	84	4937	2.618	ug/l	93
25) 2-Butanone	4.574	43	15038	10.929	ug/l	97
37) Ethyl Acetate	4.727	43	5276	1.802	ug/l #	93
43) Isopropyl Acetate	6.342	43	110022	21.826	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080423\
Data File : VX036820.D
Acq On : 04 Aug 2023 19:07
Operator : JC/MD
Sample : 03887-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MCLEAN-TP7

Quant Time: Aug 04 23:10:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

