

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080425\
 Data File : VX047217.D
 Acq On : 04 Aug 2025 10:11
 Operator : JC/MD
 Sample : VX0804MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0804MBL01

Quant Time: Aug 05 02:53:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	246208	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	453042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	439424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	224628	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156565	51.049	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.100%
35) Dibromofluoromethane	5.391	113	131220	46.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.820%
50) Toluene-d8	8.647	98	470108	51.895	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.800%
62) 4-Bromofluorobenzene	11.079	95	199023	50.982	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080425\
Data File : VX047217.D
Acq On : 04 Aug 2025 10:11
Operator : JC/MD
Sample : VX0804MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0804MBL01

Quant Time: Aug 05 02:53:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

