

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080425\
 Data File : VX047218.D
 Acq On : 04 Aug 2025 10:37
 Operator : JC/MD
 Sample : VX0804WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0804WBL01

Quant Time: Aug 05 02:54:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	365562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	684835	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	656222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	335604	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	250060	54.913	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.820%
35) Dibromofluoromethane	5.391	113	210230	49.719	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.440%
50) Toluene-d8	8.647	98	746150	54.489	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.980%
62) 4-Bromofluorobenzene	11.079	95	315122	53.400	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080425\
Data File : VX047218.D
Acq On : 04 Aug 2025 10:37
Operator : JC/MD
Sample : VX0804WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0804WBL01

Quant Time: Aug 05 02:54:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

