

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080518\
 Data File : VX003856.D
 Acq On : 03 Aug 2018 18:41
 Operator : JC/MD
 Sample : J4316-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T40-STONE-1

Quant Time: Aug 04 00:57:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080318W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 03 01:55:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	79209	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	116901	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	111810	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	60685	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	51923	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
35) Dibromofluoromethane	5.51	113	43039	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
50) Toluene-d8	8.72	98	168292	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	11.14	95	58387	45.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.82%	
Target Compounds						
11) Tert butyl alcohol	3.07	59	1661	7.34	ug/l	96
16) Acetone	2.45	43	5763	10.49	ug/l	97
20) Methylene Chloride	2.86	84	9498	9.82	ug/l	98
25) 2-Butanone	4.71	43	2147	2.30	ug/l	99
43) Isopropyl Acetate	6.47	43	42030	19.21	ug/l	99
74) N-amyl acetate	6.47	43	42030	21.78	ug/l #	100
95) Naphthalene	13.83	128	9586	2.20	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080518\
Data File : VX003856.D
Acq On : 03 Aug 2018 18:41
Operator : JC/MD
Sample : J4316-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T40-STONE-1

Quant Time: Aug 04 00:57:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080318W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 03 01:55:28 2018
Response via : Initial Calibration

