


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080518\
Data File : VX003846.D
Acq On : 03 Aug 2018 10:47
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080318W.M
Title : SW846 8260
Last Update : Fri Aug 03 01:55:28 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	1518	PASS
75	95	30	60	45.1	4281	PASS
95	95	100	100	100.0	9493	PASS
96	95	5	9	7.1	676	PASS
173	174	0.00	2	1.1	102	PASS
174	95	50	100	95.5	9068	PASS
175	174	5	9	7.0	637	PASS
176	174	95	101	98.9	8965	PASS
177	176	5	9	7.0	632	PASS